

Type Approval No : H-06-04-4

Certificate of Type Approval

1. Manufacturer : GTEC INDUSTRY INC
2. Representative : Kim, Dong-Kwan
3. Address : 1320-10, Sindang-Dong, Dalseo-Gu, Daegu
Korea.
4. Type of meter : Diaphragm gas meter (G4)
5. Identification of certified type :
 - Maximum flowrate (Q_{max}) : 6 m³/h
 - Transitional flowrate (Q_t) : 1.2 m³/h
 - Minimum flowrate (Q_{min}) : 0.04 m³/h
 - Maximum Working pressure (P_{max}) : 50 kPa
 - Cycle volume (V) : 1.3 dm³
 - Accuracy Class : 1.5
 - The range of metering temperatures (t_m) : -20°C ~ 60°C
 - Nominal internal diameter of the connecting pieces : M34×1.5
 - Remote indicating device : Yes

This is to certify that the type of measuring instrument has been certified in accordance with the provision of Clause 1., Article 12 of the Law in Metrology and Clause 1., Article 13 of the enforcement decree of the said law.

23 September 2011



President

Korea Testing Certification
Republic of Korea



Test Report of Certificate

Receipt NO : 2011-003

Issuance Date : 2011 - 09 - 23

① Manufacturer	GTEC INDUSTRY INC		② Telephone NO	053-585-2401	
③ Representative	Kim, Dong-Kwan		④ Enterprise NO	503-81-69467	
⑤ Address	1320-10, Sindang-Dong, Dalseo-Gu, Daegu Korea				
⑥ Type of Meter	Gas-Meter	⑦ Model	Diaphragm Gas Meter G4		
⑧ Application Type	New	⑨ Application Date	2011. 09. 15		
⑩ Test Contents	⑪ Oiml R137		⑫ Test Refort	⑬ Result	⑭ Remarks
- Omit the test report -				OK	

This test report that the type of measuring instrument has been in accordance with the provision of Clause 1., Article 12 of the law in Metrology and Clause 1., Article 13 of the enforcement of said Law.

Reporter : Park Ho Yong

Permission : Gwon Young Mun

Korea Testing Certification Republic of Korea



